

Date: **21 July 2026**

No.	Requirement Parameter	Specific Requirements
1	Name/Type, Purpose	Refractory porcelain crucibles. Purpose: For acid dissolution of assay beads during fire assay analysis of gold and silver (Au-Ag).
2	Place of Use	Kyrgyz Republic, Issyk-Kul Region, Kumtor Mine, Inspection and Analytical Laboratory Department (IAIL).
3	Technical Requirements	Material: High-temperature porcelain, glazed medium-form crucibles. Height: 28-30 mm. Top diameter: 35 mm. Temperature resistance: not less than 1000°C. Color: Pure white. Model: L79MF/7 (or equivalent meeting all specified parameters).
4	Supplier Requirements / Documentation to Accompany Delivery	Quality Certificate.
5	Other Requirements	Shape: Standard laboratory porcelain crucibles (cups), compatible with muffle furnaces. Structure: Uniform, free from cracks and chips.
6	Required Quantity	2,000 pcs.
7	Payment Terms and Delivery/Performance Period	To be agreed and specified in the Contract.